

Carbide Metric Wear Block Sets

Series 516 - Carbide Wear Block sets



Inspection Certificate



Carbide 2-block set




Protection gauge blocks (highlighted in grey above) are placed at each end of the gauge block stack to absorb the wear caused by contact with workpieces.

Blocks per Set	No.	Block allocation			
		Grade	Size	Step	Quantity
2	516-807-10	0	1	-	2
	516-806-10*	1			
2	516-803-10*	0	2	-	2
	516-802-10	1			

Specifications

Accuracy

EN ISO 3650



Mitutoyo Gauge Blocks and Inspection Certificates

A Inspection Certificate is furnished with all Mitutoyo gauge blocks with a serial number on the box (in the case of sets) and an identification number on each block. The deviation of each block from nominal length, at the time of inspection, is stated. For this inspection, each gauge block is measured relative to the upper level master using a gauge block comparator. Grade K gauge blocks are measured by a primary measurement method using an interferometer.