

Steel Micrometer Inspection Gauge Block Sets

Series 516

Specifications

Accuracy

EN ISO 3650



516-581-10



Micro checker
(See separate entry later in this chapter for a detailed description)



Calibration certificate



Inspection certificate

With calibration certificate

Blocks per Set	No.	Grade	Calibration/Standard applications	Block allocation
10	516-580-60	0	For Micrometer with 2 mm spindle pitch. (QuantuMike)	2.2, 4.8, 7.8, 10.4, 12, 15.2, 17.4, 19.6, 22.6, 25
	516-581-60	1		
	516-582-60	2		
10	516-106-60*	0	EN ISO 3611	2.5, 5.1, 7.7, 10.3, 12.9, 15, 17.6, 20.2, 22.8, 25 mm, Optical parallel (thickness = 12 mm)
	516-107-60	1	DIN 863 (1999)	
	516-108-60*	2	VDI/VDE/DGQ 2618	
10	516-135-60*	0	EN ISO 3611	2.5, 5.1, 7.7, 10.3, 12.9, 15, 17.6, 20.2, 22.8, 25 mm, Micro Checker, Optical parallel (thickness = 12 mm)
	516-136-60*	1	DIN 863 (1999)	
	516-137-60*	2	VDI/VDE/DGQ 2618	

With inspection certificate

Blocks per Set	No.	Grade	Calibration/Standard applications	Block allocation
10	516-580-10	0	For Micrometer with 2 mm spindle pitch. (QuantuMike)	2.2, 4.8, 7.8, 10.4, 12, 15.2, 17.4, 19.6, 22.6, 25
	516-581-10	1		
	516-582-10	2		
10	516-106-10	0	EN ISO 3611	2.5, 5.1, 7.7, 10.3, 12.9, 15, 17.6, 20.2, 22.8, 25 mm, Optical parallel (thickness = 12 mm)
	516-107-10	1	DIN 863 (1999)	
	516-108-10	2	VDI/VDE/DGQ 2618	
10	516-135-10*	0	EN ISO 3611	2.5, 5.1, 7.7, 10.3, 12.9, 15, 17.6, 20.2, 22.8, 25 mm, Micro Checker, Optical parallel (thickness = 12 mm)
	516-136-10*	1	DIN 863 (1999)	
	516-137-10*	2	VDI/VDE/DGQ 2618	

Steel Caliper Inspection Gauge Block Sets

Series 516



516-526



516-124



Inspection certificate

With inspection certificate

Blocks per Set	No.	Grade	Calibration/Standard applications	Block allocation
4	516-526-10	1	EN ISO 13385-1	4 blocks : 10, 30, 50, 125 mm, Setting ring (Ø4 mm, Ø10 mm), Pin gauge (Ø 10 mm), Glove
	516-527-10*	2		
3	516-124-10	1	DIN 862 (1988)	3 blocks : 30, 41.3, 131.4 mm, Setting ring (Ø4 mm, Ø25 mm), Glove
	516-125-10	2		